

SEP 26 2011

SEP 26 2011

VERIFICATION OF CALIBRATION REPORT*of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N <u>130176</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>DARYL RICE</u>	ID# <u>2088</u> Date <u>09/22/2011</u>
A Agency <u>SITKA POLICE DEPT.</u>	Phone # <u>747-3245</u>
Instrument Location <u>304 LAKE STREET</u>	
B Alco S/N <u>X172969</u> Target Value <u>.077</u> High Pressure <u>800</u>	
Alco Test Values	<u>.079</u> 1st Alco
	<u>.079</u> 2nd Alco
Signature <u>Daryl Rice</u>	<u>COB</u> <u>9/26/11</u>
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Nov, 2010.

Nikki Roth
(Notary Seal Stamp)
Daniel L. Nicholas, Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

SEPTEMBER 22, 2011

OPERATOR'S NAME:

RICE DARYL DEAN

OPERATOR'S NUMBER: 2888

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: SIT1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .077

ALCO S/N: X172969

— BREATH ANALYSIS —

.077 ADJUSTED FOR 29.90 in
ALCO TARGET .076 07:18
BLANK TEST .000 07:20
INTERNAL STANDARD VERIFIED 07:20
ALCO TV 29.90 in .079 07:20
BLANK TEST .000 07:21
SUBJECT SAMPLE .000 07:21
BLANK TEST .000 07:22
ALCO TV 29.90 in .079 07:22
BLANK TEST .000 07:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

SEPTEMBER 22, 2011

TIME 07:25

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~